

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|--|------------------|---------------------------|----------|
| Item No.: | 1 | Quantity | 6 |
| ABC : | P6,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">FOOT STOOL</p> <p>Product Description</p> <ul style="list-style-type: none"> • Flared legs with rubber tips. • All-welded construction. • Single-step, with non-slip rubber mat. • Stainless steel. <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p style="text-align: center;">Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|----------|---------------------------------|----------|
| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 2 | Quantity | 2 |
| ABC : P7,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>CARDIAC BOARD Size: Manufacturer Default TECHNICAL FEATURES: * The CPR Board provides a firm surface, elasticity. * Help to fix the position of patient and ensure the patient open airway. * The head cup maintains proper alignment with the patient's head. * The board adopts PE material with no discharge contaminator. Standard Requirements: ·Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate Warranty Certificate ·One (1) Year for craftsmanship, after inspection and acceptance Delivery 60 days upon receipt of Purchase Order Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable ***** NOTHING FOLLOWS *****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME

TELEPHONE / FAX NO.

OF AUTHORIZED REPRESENTATIVE

Technical Specifications

| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
|--|---|--------------------------|---|
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 3 | Quantity | 3 |
| ABC : P9,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>THERMOMETER TYMPHANIC TECHNICAL SPECIFICATIONS: Type: Ear Thermometer 1. Measurement Range: 89.6°F ~ 108°F (32°C ~ 42.2°C) "2. Measurement Accuracy: ±0.2 °C, 32.0 ~ 42.2 °C ±0.4 °F, 89.6 ~ 108.0 °F" 3. LCD or LED temperature reading display in Celsius scale 4. Battery operated 5. Components must have no glass and mercury materials 6. Beep sound upon measurement 7. With low battery indicator 8. The battery must provide a minimum of 1000 measurements before low battery indicator appears on display Mobility / Portability: Handheld Device Standard: Device corresponds to the requirements of the standard for infrared thermometers ASTM E 1965, EN 12470-5, IEC 60601-1, IEC 60601-1-2 Electromagnetic: Device fulfills the stipulations of the standard Compatibility: IEC 60601-1-2 Standard Requirements: • Delivery ___days upon receipt of Purchase Order</p> | | | |

- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 4 | Quantity | 2 |
| ABC : P12,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>PATIENT CHART HOLDER (12 CAPACITY) Made of Plain Sheet G-22 Chart file 12 capacity Cabinet type with Drawer & Rubber Bumper. Mounted on four wheels' caster White paint or stainless Size L: 22" x H: 33.5" x W: 23"</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery ____days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p style="text-align: center;">Warranty Certificate</p> | | | |

| | |
|---|--|
| <ul style="list-style-type: none"> • Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user <p style="text-align: center;">Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*****Nothing Follows*****</p> | |
|---|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-------------------|---------------------------|-----------|
| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 5 | Quantity | 28 |
| ABC : | P14,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |

PATIENT CHART ALUMINUM

Size Long: (14 x 9")

Standard Requirements:

- Delivery period _____ calendar days.
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|------------------|----------|-----------------|-----------|
| Item No.: | 6 | Quantity | 15 |
|------------------|----------|-----------------|-----------|

| | |
|---|---------------------------------|
| ABC : P15,000.00 | Country of Origin: |
| Brand: | Model: |
| PURCHASER'S SPECIFICATION | SUPPLIER'S SPECIFICATION |
| <p>NEUROLOGICAL HAMMER</p> <p>Chrome-plated octagonal handle</p> <ul style="list-style-type: none"> • Compartment for stainless steel needle and brush • Black rubber bumper • Overall length: 7" <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> • The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility • Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> • Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. • List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions:</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*****Nothing Follows*****</p> | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| |
|---|
| <p>Republic of the Philippines Department of Health</p> |
|---|

CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|--|----------|---------------------------------|----------|
| Item No.: | 7 | Quantity | 1 |
| ABC : P16,500.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">WEIGHING SCALE WITH MEASURING STICK, ADULT</p> <p>Mechanical weighing scale with height rod 220kgs maximum weighing capacity Non-slip steel platform Adjustable/telescopic height rod with 1mm graduation Engraved numbers on eye level scale beam and height rod Large low rise platform (at least 13 inches' length x 13 inches' x 3 inches' high Weight increments/graduation of 50grams Height rod of aluminum material Powder coated body</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery _____days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p align="center">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 8 | Quantity | 4 |
| ABC : P18,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">Examining Lights/Drop Lights</p> <p>Specifications:</p> <ul style="list-style-type: none"> • Flexible light head • Power supply: 220 VAC, 60Hz • Type: Mobile <p>Adjustable height: at least 400mm</p> <p>Material:</p> <ul style="list-style-type: none"> • Aluminum Cast Alloy Five Stroke Base Castor with at least two (2) Castor Locks/Support <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|------------|--------------------------|---|
| Item No.: | 9 | Quantity | 1 |
| ABC : | P20,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| LINEAR MICROPIPETTE/PIPETTOR STAND AND RACK, HOLDS 6 PIPETTES Product description: 6 PIPETTORS IVYX Scientific Micropipettes are general purpose micropipettes for the accurate and precise sampling and dispensing of liquid volumes. They operate on the air displacement principle and use universal-size disposable tips. All pipettes have been quality tested according to ISO8655/DIN12650. The quality control according to ISO ISO8655/DIN12650 involves gravimetric testing of each micropipette with distilled water (quality 3, DIN ISO 3696) at 22°C using the manufacturer's original tips. Each box includes 1. IVYX Scientific Micropipette 2. Calibration tool. 3. Removal tool (to remove the bottom half for cleaning or autoclaving. 4. User guide. 5. Micropipette holder. 6. One tip. 7. Quality control certificate. 8. Small vial of silicon grease (use after cleaning) PIPETTE STAND A sturdy pipette stand, capable of holding 6 micropipettes. PIPETTE TIPS | | | |

Five racks of 96 pipette tips each. RNase and DNase free, sterilized by Electron Beam radiation. 10µl size rack fits the 0.1-2.5µl and 0.5-10µl pipettors. 200µl size rack fits the 10-100µl pipettor, 1000µl size rack fits the 100-1000µl pipettor, 5000µl size rack fits the 1000-5000µl pipettor, and 10ml size rack fits the 2-10ml pipettor

Standard Requirements:

- Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate Warranty Certificate
- One (1) Year for craftsmanship, after inspection and acceptance

Delivery 60 days upon receipt of Purchase Order

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

***** NOTHING FOLLOWS *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 10 | Quantity | 9 |
| ABC : P45,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">MANUAL RESUSCITATOR PEDIA</p> <p>Standard Features</p> <ul style="list-style-type: none"> made of 100% silicon with visible markings in the bag as autoclavable all parts can be easily removed to ease cleaning or disinfecting with pressure relief valve Capacity: 650 to 750ml for pedia <p>Accessories</p> <ul style="list-style-type: none"> face mask - 2 pcs oral airway - 3 pcs reservoir bag and oxygen tubing should be in a plastic container <p>Standard Requirements:</p> <ul style="list-style-type: none"> Delivery period _____ calendar days. Current and Valid Certificate of Manufacturer's compliance with ISO 13485 Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market | | | |

Warranty Certificate

- Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 11 | Quantity | 19 |
| ABC : P47,500.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>BP APPARATUS WITH ADULT AND PEDIA CUFF, WALL TYPE</p> <p>Features:</p> <ul style="list-style-type: none"> • Aneroid • Pressure graduation 0 up to 300mmHg <p>Standard Accessories</p> <ul style="list-style-type: none"> • Bandage cuff (cotton made) • Bladder • Rubber • Bulb with airflow control • Cuff size: adult • Latex free rubber • Bulb <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery _____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier | | | |

- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

- Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

| TECHNICAL SPECIFICATIONS | | | |
|---|----|--------------------------|---|
| Item No.: | 12 | Quantity | 3 |
| ABC : P51,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">DRESSING CART</p> <ul style="list-style-type: none"> • Four pieces' stainless jars with covers • One stainless forceps jar • Eight pieces' reagent bottles with stopper • Stainless pail, bowl and tray, two drawers, shelves with railings, push handle, mounted on heavy duty casters • Dimension at least: W-25" L-17 ¾" H-30 ¼" <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery _____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p align="center">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 13 | Quantity | 13 |
| ABC : P65,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">MANUAL RESUSCITATOR, ADULT</p> <p>TECHNICAL SPECIFICATIONS</p> <ol style="list-style-type: none"> 1. completely autoclavable, heavy duty 2. made of 100% silicon with visible markings in the bag as autoclavable 3. all parts can be easily removed to ease cleaning or disinfecting 4. with pressure relief valve 5. with 2 pieces face mask (pedia & adult), 3 pieces' oral airway (infant, pedia & adult), reservoir bag and oxygen tubing 6. capacity: 1600ml for adult 7. resuscitator can be totally disassembled, easy to clean and disinfect 8. all parts must be of high-strength, long-life materials and require no special maintenance or storage conditions 9. resuscitator must be supplied as a complete set with: <ol style="list-style-type: none"> a. non-rebreathing patient valve with pressure limiting valve b. compressible self-refilling ventilation bag with an approximate capacity of at least 250ml c. intake valve with nipple for O2 tubing d. oxygen reservoir bag complete with approximate capacity of 2 liters e. masks: adult size in 2 different sizes, translucent and round type | | | |

| | |
|---|--|
| <p>f. airways Guedel: adult size in 2 different sizes, translucent</p> <p>10. materials:</p> <p>a. non-rebreathing patient valve with pressure limiting valve: polycarbonate/ polysulfone</p> <p>b. compressible self-refilling ventilation bag: silicon rubber</p> <p>c. intake valve with nipple for O2 tubing: polycarbonate/polysulfone</p> <p>d. oxygen reservoir bag: translucent plastic</p> <p>e. masks: silicone rubber</p> <p>f. airways Guedel: plastic</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p> | |
|---|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|-------------------|---------------------------|----------|
| Item No.: | 14 | Quantity | 7 |
| ABC : | P70,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>MINOR SURGICAL SET/CUTDOWN SET</p> <p>Materials: Stainless steel, surgical grade in accordance with ISO standard</p> <p>Components: 1 pc – Surgical knife handle #3 (Scalpel) 1 pc – Mayo heager needle holder, bulldog jaw, 16cm (6-1/4") 1 pc – Instrument tray with cover, tarnished, 18/8, stainless steel (410mm x 254mm x 64mm) 1 pc – Tissue forceps 1 x 2 teeth, 5-3/4" (14.5cm) 4 pcs – Halsted-Mosquito Haemostatic forceps straight 5" (12.5cm) 20mm jaw with full serrations 4 pcs – Micro-Mosquito Haemostatic forceps, curved, 4-3/4" (12cm) 20mm jaw with full serrations 1 pc – Mayo Operating scissors 5-3/4" straight (14.5cm)</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery | | | |

- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 15 | Quantity | 1 |
| ABC : P70,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">INSTRUMENTS FOR SPINE SURGERY SPINAL PEDICLE SCREW SYSTEM INSTRUMENT SET</p> <p>Material: Stainless Steel</p> <p>Usage: Spinal Fixation</p> <p>Specification: Feeler for Screw Channel Straight #1 Hex Screwdriver for Srew Nut SW5.0 #1 Tap ϕ4.0 #1 Tap ϕ5.0 #1 Screwdriver Hex 3.5mm for Crosslink Nut #1 Tap ϕ6.0 #1 Tap ϕ7.0 #1 Feeler for Screw Channel Bent #1 Screw Cutter for Long Arm Screw #1 Screw Nut Holder Hex #1 Crosslink Nut Holder Hex #1 In-situ Bending Iron #1 Screwdriver for Polyaxial Screw #1 Screwdriver for Monoaxial Screw #1 Trail Rod #1 Insert Device for Fixation Pin #1 Rod Pusherial #1 Parallel Spreader #1</p> | | | |

| | |
|--|--|
| Parallel Compressor #1 Rod Twist #1 Pedicule Probe Straight #1 Pedicule Probe Bent #1 AWL #1 T-handle Quick Coupling #1 Rod Holding Forcep #1 Rod Bender #1 Counter Torque for Screw Cutter #1 Rod Pushing Forcep #1 Straight Handle Quick Coupling #1 Counter Torque #1 Aluminium Box #1 Standard Requirements: ·Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate Warranty Certificate ·One (1) Year for craftsmanship, after inspection and acceptance Delivery ____ days upon receipt of Purchase Order Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable ***** NOTHING FOLLOWS ***** | |
|--|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 16 | Quantity | 18 |
| ABC : P72,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">NEBULIZER (HEAVY DUTY)</p> <p>Technical Specifications Power Supply: 220-240 VAC , 60Hz Standard Feature: Compressor (flapper type) Standard Accessories: 2 pcs Adult mask 2 pcs Child mask 2 sets Sterilizable tubings 2 pcs. air Filter 2 pcs Mouthpiece (disposable) ABS or hard plastic body Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery ____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the | | | |

account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery

- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 17 | Quantity | 14 |
| ABC : P84,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">STETHOSCOPE, PEDIA & NEONATE</p> <p>Standard Features:</p> <ul style="list-style-type: none"> • Double-sided Chest piece (Wt:1 - 1.5 ounce) • Machined Stainless Steel Chest Piece • Diaphragm made with epoxy or fiberglass (Dm: 2.5 - 3.3 cm) • Soft sealing rubberized ear tips • Rubberized tubing's • Head set wide diameter aerospace alloy • Length: 70 - 75 cm • Weight: 100 - 110 gm. <p>Standard Accessories:</p> | | | |

- 1 pair Spare Ear tips
- Spare Diaphragm

Standard Requirements:

- Delivery ____days upon receipt of Purchase Order
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|------------|--------------------------|----|
| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 18 | Quantity | 31 |
| ABC : | P93,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |

**THERMOMETER(INFRARED-NON
CONTACT)**

Specification:

- *Temperature range:-50 - 330°C (-58-626°F)
- **Accuracy:±1.5% or ±1.5°C
- *Repeatability:±1% or ±1
- *Distance Spot Ratio:12:1
- *Emissivity:0.95 preset
- *Resolution:0.1°C/0.1°F
- *Response Time:500ms
- *Wavelength:8-14µm
- *°C/°F Selection
- *Data Hold function
- *Laser Target Pointer ON/OFF selection
- *Backlight ON/OFF selection
- *Auto Power Shut Off in 7 seconds without any operation
- *Power supply:1.5V battery
- *Weight:147.5g(Including battery)
- *Dimension:153*101*43mm

Standard Requirements:

- Delivery _____ days upon receipt of Purchase Order
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|-----------|----|----------|---|
| Item No.: | 19 | Quantity | 1 |
|-----------|----|----------|---|

| | |
|--|---------------------------------|
| ABC : P100,000.00 | Country of Origin: |
| Brand: | Model: |
| PURCHASER'S SPECIFICATION | SUPPLIER'S SPECIFICATION |
| <p>CLINICAL CENTRIFUGE 24 PLACER</p> <p>Capacity: 24 placer (24 X 5 – 10 ml test tubes)</p> <p>Power Supply Input: 220 – 240 VAC, 60Hz</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> • Brushless motor drive • max. speed: 4,000 RPM • Fixed Rotor • Lid dropping protection, with metal lid • Metal housing, lid locking and holding during rotor run • One hand lid lock; emergency lid lock release • Stainless steel centrifuging chamber, air cooling • Automatic rotor recognition • Error display, imbalance detection <p>Standard Accessories:</p> <ul style="list-style-type: none"> • Automatic voltage regulator (compatible to the equipment power requirement) <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 and CE Certificate • Users Manual in English Language <p>a. Operational Manual - 2 copies</p> <p>b. Service Manual – 2 copies</p> <ul style="list-style-type: none"> • Brand must be in the local and international market for at least 10 years <p>Training requirements</p> <ul style="list-style-type: none"> • Certification that winning bidder will provide training for at least two (2) end user (Medical Technologist and Equipment Maintenance). • Training Certificate must be noted/signed by the Head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> • Certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the warranty period. • List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available. <p>Warranty Certificate</p> <p>Warranty Certificate of One (1) year for parts and two (2) years for service including calibration and preventive maintenance, after the provision of the required training of the end user.</p> <p>*****Nothing Follows*****</p> | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|-----------|---------------------------|----------|
| Item No.: | 20 | Quantity | 1 |
| ABC : P100,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| MICROCENTRIFUGE 24 PLACER Description: Fast acceleration to a maximum speed 13,300 rpm/16,300 x g Complete with easy access, 24 x 1.5/2.0 mL rotor; Precise, digital control system Set and view rotor speed in rpm or g-force Quiet and cool running, even at maximum speed SPECIFICATIONS Net Weight: 7.8 lb. (8.1kg) Max. Speed: 13,300rpm Display: LED Height (Metric) Exterior: 21.6 cm Length (Metric) Exterior: 23.5 cm Width (Metric) Exterior: 29.3 cm Maximum Force Range: 16,000 to 16,999 x G Refrigerated: No Capacity 24 x 1.5/2 MI Max. RCF: 16,999 x G Height (English) Exterior: 8.5 in. Length (English) Exterior: 9.2 in. Width (English) Exterior: 11.5 in. Electrical Requirements: 120V 60Hz Rotor Included: Yes Standard Requirements: ·Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate Warranty Certificate ·One (1) Year for craftsmanship, after inspection and acceptance Delivery ____ days upon receipt of Purchase Order Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable ***** NOTHING FOLLOWS ***** | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|-----------|---------------------------|----------|
| Item No.: | 21 | Quantity | 9 |
| ABC : P117,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>REFRIGERATOR SINGLE DOOR 6.5 CU FT.</p> <p>Type: Single Door</p> <p>Size:</p> <p>Dimension: 1234 x 525 x 560 (mm)</p> <p>Capacity:</p> <p>Capacity (Total): 149.3L</p> <p>Freezer: 25L</p> <p>Refrigerator: 119L</p> <p>Other: 5.3L</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period ____ calendar days. • Certification to provide user's Manual in English Language <ul style="list-style-type: none"> a. Operations Manual 2 copies b. Service Manual 2 copies • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> • The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility • Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> | | | |

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

Warranty Certificate

Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*** Nothing Follows ***

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 22 | Quantity | 25 |
| ABC : P125,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">STETHOSCOPE ADULT</p> <p>Standard Features:</p> <ul style="list-style-type: none"> • Single-sided Chest piece (Wt: 3 - 4 ounce) • Black, Plated and Polished alloy Chest Piece • Tunable diaphragm made with epoxy or fiberglass (Dm: 1.5 - 2 inches) • Soft sealing rubberized ear tips • Rubberized tubings • Head set wide diameter aerospace alloy • Length: 25 - 30 inches • Weight: 150 - 170 gm. <p>Standard Accessories:</p> <ul style="list-style-type: none"> • 1 pair Spare Ear tips • Spare Diaphragm <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery ____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> | | | |

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and
acceptance and after the provision of the required
training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously
displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|-----------|
| <p align="center"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| <p align="center">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 23 | Quantity | 16 |
| ABC : P128,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">WHEEL CHAIR (I.V STAND)</p> <p>Standard Features Heavy duty, locally made, stainless steel or chrome plated, durable Foldable Leatherette or equivalent seat and back rest foot rest wheel lock handle</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery ____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery | | | |

- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|----|--------------------------|----|
| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 24 | Quantity | 24 |
| ABC : P132,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |

**BP APPARATUS, STAND TYPE, ANEROID
(WITH ADULT & PEDIA CUFF)**

Standard Accessories:

- With Stand

Standard Accessories:

- Bandage cuff (adult and pedia) (Velcro type)
- Inflation bag (adult and pedia) (latex free)
- Rubber valve with airflow control (latex free)
- Preferably made in US or Germany

Standard Requirements:

- Delivery ____ days upon receipt of Purchase Order
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health

CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|-----------|---------------------------|-----------|
| Item No.: | 25 | Quantity | 45 |
| ABC : P135,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">EMERGENCY LIGHT (PORTABLE)</p> <p>Product Description</p> <p>Standard Features</p> <ul style="list-style-type: none"> • LED Lamp • Power Supply: 220 VAC/12 VDC • Battery: 6V/4Ah Sealed Lead-acid Free • At least 6 operating hours (fully charged battery) <p>Maintenance Rechargeable Battery</p> <p>Charging Indicator will be Red and Turned into Green at Full Charge</p> <p>Built in Overcharge Circuit</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p align="center">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 26 | Quantity | 1 |
| ABC : P450,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>BIOLOGICAL SAFETY CABINET 4 FT.</p> <ul style="list-style-type: none"> • Imported from Germany, Japan, Italy, USA, Israel/ France, Korea • Class II, Type A • Wind Flow type: Recirculation 70 %, Exhaust 30% • Internal Dimension: 1340 x 623 x 670 mm (approximate) • Sterilization: U.V. 20 W • Material (Interior): Stainless; (Exterior): Powder Coating • with HEPA Filter, High glass Fiber Type Filter, 0.3 um, • Particle 99.97% Cut-off or better • To provide stand/frame • High efficiency double bower motor with double dustproof structure for low noise and low vibration <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate <p>Warranty Certificate</p> <ul style="list-style-type: none"> • Preventive maintenance every quarter w/in the warranty • One (1) Year for craftsmanship, after inspection and acceptance • Delivery ____days upon receipt of Purchase Order <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">***** NOTHING FOLLOWS *****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|---|-------------|--------------------------|----|
| Item No.: | 27 | Quantity | 60 |
| ABC : | P180,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>OXYGEN TANK CARRIER (SINGLE)</p> <p>Materials:</p> <ul style="list-style-type: none">• Tubular Steel frame (Enamel Coated) <p>Features</p> <ul style="list-style-type: none">• tank holder (Single Tank, 50 Lbs.)• Two heavy duty casters <p>Standard Requirements:</p> <ul style="list-style-type: none">• Delivery period _____ calendar days.• Current and Valid Certificate of Manufacturer's compliance with ISO 13485• Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier• Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery.• The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> | | | |

| | |
|--|--|
| 2. Property tagging (sticker) rewritable ***** Nothing Follows ***** | |
|--|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|--------------------------|----------|
| <p align="center"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| <p align="center">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 28 | Quantity | 8 |
| ABC : P200,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |

EENT DIAGNOSTIC SET

Technical Specifications

- Stainless Steel

Includes the following instruments:

- Otoscope, with two levels of magnification
- Ophthalmoscope, with different levels of magnification
- Stainless steel battery handle with brightness control
- Nasal speculum
- Bent arm throat lamp
- Illuminated tongue depressor
- Post-nasal mirror 18mm diameter
- Laryngeal mirror 22mm diameter
- 4 pcs. Rechargeable battery with charger of 220V, 60 Hz
- Spare light bulb (LED/Xenon)

Standard Requirements:

- Delivery period _____ calendar days.
- Certification to provide users Manual in English Language
 - a. Operations Manual 2 copies
 - b. Service Manual 2 copies
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Training Requirement

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

After Sales Service

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.

- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

Warranty Certificate

Minimum of three (3) years for both parts & service
Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*** **Nothing Follows** ***

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|--|--------------------|---------------------------------|----------|
| Item No.: | 29 | Quantity | 4 |
| ABC : | P320,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">SYRINGE PUMP</p> <p>Instrument classification: Class II Weight: 1.7KG Open System: Universal Syringes, edit & store no less than 20 brands Adjustable Flow Rate Range: 0.1-1500ml/h, min. increment 0.01ml/h Display: 4.3 inches' full color touch screen Infusion Modes: 3 infusion modes Alarm system: Visible & audible with color, text and sign indication System Language: 12 languages, support customize Power supply: AC, DC & Battery Power Source: AC, DC and battery Standard Requirements: <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate Warranty Certificate <ul style="list-style-type: none"> • One (1) Year for craftsmanship, after inspection and acceptance • Delivery __ days upon receipt of Purchase Order Special Provisions <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">***** NOTHING FOLLOWS *****</p> </p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 30 | Quantity | 2 |
| ABC : P280,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">High Speed Sterilizer (Table top)</p> <p>Type: Water Filter</p> <p>Feature: Without Bag</p> <p>Application: Dry</p> <p>Features:</p> <ol style="list-style-type: none"> 1. The autoclave strictly meets the standard of EN13060, 93/42/CE and adopts European B standard 2. Super-fast speed of the whole process, only 6 minutes 3. Five times per-vacuum preceding vacuum drying, the vacuum measurement reaches-0.098Mpa and residual humidity is less than 0.2%, suitable for miscellaneous endoscopes, instrument and longer pipes 4. 3-fixed, 1- customized program. 239Fahrenheit, 249.8 Fahrenheit, 273.2 Fahrenheit. The sterilization time and vacuum time can be adjusted by manual, ensure user obtain best sterilization effects according to private requirements 5. Microprocessor controlled. Malfunction detection system, corresponding error code will be showed on the display once there's something goes wrong 6. Having passed B/D test, the leak rate of vacuum is less than 0.13Kpa/min 7. Double safety locking device prevents door from opening while chamber is pressurized 8. An automatic safety shut off for self-protection in case of excessive temperature and pressure <p>Volume : 10liter</p> <p>Voltage Power: 220V/55Hz</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate <p>Warranty Certificate</p> <ul style="list-style-type: none"> • One (1) Year for craftsmanship, after inspection and acceptance | | | |

| | |
|---|--|
| <ul style="list-style-type: none">• Delivery ____ days upon receipt of Purchase Order <p>Special Provisions</p> <ol style="list-style-type: none">1. Non-removable DOH letters conspicuously displayed on the front2. Property tagging (sticker) rewritable <p>***** NOTHING FOLLOWS *****</p> | |
|---|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 31 | Quantity | 4 |
| ABC : P300,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">BLOOD GAS AND CHEMISTRY ANALYSIS SYSTEM/PORTABLE BLOOD GAS ANALYZER</p> <p>Throughput: Results in 1 minute after sample aspiration</p> <p>Sample volume: 140µl</p> <p>Quality control: 3 or 5 levels QC, External electronic simulator</p> <p>Display: 7-inch color LCD Display</p> <p>Interface: 4 x USB 2.0 host, 1 x RS232, WLAN</p> <p>Input device: Touch screen and barcode scanner</p> <p>Power supply: 100~240VAC, 50/60Hz</p> <p>Battery: 5000mAh rechargeable lithium-ion battery, 50 samples continuous testing</p> <p>Weight: 3.65kg</p> <p>Operation Environment: 10 °C-31 °C; %RH: 25%-80%; 70 -106.6 KPa</p> <p>Parameter -Reportable Range</p> <p>pH (pH units) - 6.5 – 8.0</p> <p>pO₂ (mmHg)-10 – 700</p> <p>pCO₂ (mmHg) -10 – 150</p> <p>Na (mmol/L) -100 – 180</p> <p>K (mmol/L)-2.0 – 9.0</p> <p>Ca (mmol/L)-0.25 – 2.5</p> <p>Cl (mmol/L)-65 – 140</p> <p>Glu (mg/dL or mmol/L)-20 – 700/1.1 – 38.9</p> <p>Lac (mg/dL or mmol/L)-2.7 – 180.2/0.3 -20</p> <p>Hct (%PCV or Fraction)- 10 – 75 /0.10 – 0.75</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate <p>Warranty Certificate</p> | | | |

| | |
|--|--|
| <ul style="list-style-type: none">• One (1) Year for craftsmanship, after inspection and acceptance• Delivery _____ days upon receipt of Purchase Order <p>Special Provisions</p> <ol style="list-style-type: none">1. Non-removable DOH letters conspicuously displayed on the front2. Property tagging (sticker) rewritable <p>***** NOTHING FOLLOWS *****</p> | |
|--|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| <p style="text-align: center;">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 32 | Quantity | 9 |
| ABC : P333,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>BED, MECHANICAL 2 CRANKS WITH SIDE RAILS AND MATTRESS</p> <p>Standard Features</p> <ul style="list-style-type: none"> • Dimensions • Bed Table • Length: at least 80 inches • Width: at least 35 inches • Height: at least 18 inches • Back Section Lifting angle: 60o-65o • Leg Section Lifting Angle: at least 30o • Casters with lock-5" <p>Specific Characteristics</p> <ul style="list-style-type: none"> • 2 cranks • Steel frame Epoxy Painted • Detachable ABS head and Foot Boards • Foldable cranks handle with over-loading protection system • Support Mattress Platform: ABS Molding or metal sheet • With Side Rails (at least ½ Length) Aluminum Alloy • IV socket; 4 (one on each side) • With patient name holder at foot and head Part | | | |

Accessories

- Adjustable or retractable IV pole (4 hooks)
- Mattress 4"-5" thick rubberized foam with leatherette cover
- Over bed Table-Formica top with adjustable height
- With one step foot stool

Standard Requirements:

- Delivery period _____ calendar days.
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturer's certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Training Requirement

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

After Sales Service

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

Warranty Certificate

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
|---|----|--------------------------|---|
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 33 | Quantity | 3 |
| ABC : P420,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p>AUTOCLAVE TABLE TOP 20 L CAPACITY Chamber capacity: 20 L Power Supply: 220VAC / 60 Hz Safety Device:</p> <ul style="list-style-type: none">• Over Heat (Low Water)• Cut off switch• safety valve and release valve• emergency exhaust valve• door safety lock• pressure gauge with pressure control times with alarm <p>Standard Accessories:</p> <ul style="list-style-type: none">• Stainless steel basket• Tong <p>Standard Requirements:</p> <ul style="list-style-type: none">• Delivery period ____ calendar days.• Certification to provide user's Manual in English Language <p>a. Operations Manual 2 copies</p> | | | |

| | |
|--|--|
| <p>b. Service Manual 2 copies</p> <ul style="list-style-type: none"> Current and Valid Certificate of Manufacturer's compliance with ISO 13485 Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> Non-removable DOH letters conspicuously displayed on the front Property tagging (sticker) rewritable <p style="text-align: center;">*** Nothing Follows ***</p> | |
|--|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|--------------------|---------------------------|-----------|
| Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 34 | Quantity | 49 |
| ABC : | P490,000.00 | Country of Origin: | |
| Brand: | Model: | | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |

| | |
|---|--|
| <p style="text-align: center;">OXYGEN TANK 50L</p> <p>TECHNICAL SPECIFICATIONS</p> <p>Oxygen Tank Cylinder:</p> <ol style="list-style-type: none"> 1. 50L capacity 2. standard regulator connector (for hospital use) 3. output control valve 4. tank control valve cover 5. tank color (green) <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery ____ days upon receipt of Purchase Order • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*****Nothing Follows*****</p> | |
|---|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health

CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

| | | | |
|--|--------------------|---------------------------------|-----------|
| Item No.: | 35 | Quantity | 14 |
| ABC : | P630,000.00 | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">EMERGENCY CART</p> <p>Product Description Size: at least 470mm x 680mm x 980mm, 250mm cabinet height steel plate body ABS Molded base and top place Pull out writing table, push handle Four drawers with ABS Plastic basket with removable dividers At least 5mm guard rubber bumper, Four (4) caster with locking brake</p> <p>Accessories: 1 pc. Stainless steel trash can</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable</p> <p align="center">*****Nothing Follows*****</p> | | | |

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|-----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 36 | Quantity | 17 |
| ABC : P680,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">LARYNGOSCOPE SET (ADULT/ PEDIA)</p> <p>Type: Battery Operated (Alkaline)</p> <p>Composition: 4 pcs – Mc Intosh blades No. 1, 2, 3, 4 2 pcs – Miller blades No. 0, 1 (6pcs Miller blades - 00,0,1,2,3,4) Laryngoscope handle with LED source</p> <p>Features: Stainless steel construction of blades Autoclavable blades</p> <p>Standard accessories:</p> <ul style="list-style-type: none"> • Lamp case • 2 pcs – LED spare lamp • Rechargeable battery with charger, 220V, 60 Hz <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery period _____ calendar days. • Current and Valid Certificate of Manufacturer's compliance with ISO 13485 • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <ul style="list-style-type: none"> • Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, | | | |

inspection and acceptance and after the provision
of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously
displayed on the front

2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|--|-----------|---------------------------|-----------|
| <p align="center">Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p> | | | |
| <p align="center">TECHNICAL SPECIFICATIONS</p> | | | |
| Item No.: | 37 | Quantity | 16 |
| ABC : P800,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p align="center">PULSE OXIMETER</p> <p>SPO2, Pulse Rate with Warning Indicators & Pletysmography Audible/ Visual Alarms for High/Low SPO2 and PR, Audible/ Visual Alarms for Low-Battery Power Requirements: 220VAC, 60Hz, Table top Transducers for adult, neonate and pedia With battery back-up for at least two (2) hours Standard Requirements:</p> <ul style="list-style-type: none"> ❖ Delivery _____ days upon receipt of Purchase Order ❖ Current and Valid Certificate of Manufacturer's compliance with ISO 13485 ❖ Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier ❖ Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery ❖ The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> | | | |

Minimum of one (1) year for craftsmanship & service
Warranty Certificate upon delivery, inspection and
acceptance and after the provision of the required
training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously
displayed on the front
2. Property tagging (sticker) rewritable

*******Nothing Follows*******

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.

Technical Specifications

| | | | |
|---|-----------|---------------------------|----------|
| <p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p> | | | |
| TECHNICAL SPECIFICATIONS | | | |
| Item No.: | 38 | Quantity | 7 |
| ABC : P700,000.00 | | Country of Origin: | |
| Brand: | | Model: | |
| PURCHASER'S SPECIFICATION | | SUPPLIER'S SPECIFICATION | |
| <p style="text-align: center;">ECG MACHINE PORTABLE</p> <p>Features</p> <ul style="list-style-type: none"> * Light in weight, compact bin size * Three kinds of operation mode: Automatic, Manual, Analysis * Standard lead mode or European / Cabera lead mode available * Automatic measurement and interpretation * Detection and alarm of lead - off, No paper, No Battery * Built in ECG simulator * Bullt in rechargeable lithium battery with high capacity | | | |

| | |
|---|--|
| <p>* Automatic adjustment of baseline for optimal recording</p> <p>* RS232 communication interface in standard configuration for connecting with PC or setting up of ECG, database, option USB and Blue tooth interface</p> <p>* Built in Help function: Electrode positioning, ECG basic knowledge, common trouble - shooting, etc.</p> <p>* Clinical information such as patient ID, Name, sex, age, height, weight, drugs and hospital name can be input</p> <p>* Case storage function: 15 cases can be saved in the machine or such as much as 150 cases in a SD card. Saved ECG waveforms can be cine-looped, printed out or transferred to PC.</p> <p>Technical Specification:</p> <p>Acquisition Mode: 12 lead acquisition simultaneously</p> <p>Input Circuit: Floating, protection againsts defibrillator and pacemaker</p> <p>A/D Converter: 18 Bit</p> <p>Time Constant: $\geq 3.2s$</p> <p>Frequency Response: 0.05Hz - 150Hz</p> <p>Sensitivity: 2.5,5,10,20 (mm/mV),Auto</p> <p>Input Impedance: $\geq 50M\Omega$</p> <p>Input Circuit Current: $\leq 50nA$</p> <p>Calibration Voltage: $1MV \pm 2\%$</p> <p>Noise Level: $< 15\mu Vp-p$</p> <p>Anti Baseline Drift: Automatic</p> <p>Filter:</p> <ul style="list-style-type: none"> - EMG Filter: 35Hz - 45Hz (-3dB) - ADS Filter: Auto (0.15.0.50Hz) - HUM Filter: 50Hz/60Hz (-20dB) <p>CMRR: $> 100dB$ (with AC filter)</p> <p>Patient Current Leakage: $< 10\mu A$</p> <p>Paper Speed: 6.25/12.5/25/50mm/s</p> <p>Recording Paper: 63mm/80mm(width), 20m/30m(length), roll/Z-fold</p> <p>Power Supply:</p> <ul style="list-style-type: none"> - AC: 110/220V,50/60Hz; 85V-265V(Option) - DC: 14.4V, built in rechargeable lithium battery <p>Dimension: 345mm x 300mm x 80mm; 415mm x 195mm x 380mm (Carton)</p> <p>Net Weight: 2.5Kgs.</p> <p>Gross Weight: 4.5Kgs.</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> • Delivery _____periodcalendar days. • Certification to provide users Manual in English Language <p>a. Operations Manual 2 copies</p> <p>b. Service Manual 2 copies</p> | |
|---|--|

| | |
|--|--|
| <p>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</p> <ul style="list-style-type: none"> • Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier • Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery • The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> • The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility • Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <p>Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.</p> <p>· List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available</p> <p>Warranty Certificate</p> <p>Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*** Nothing Follows ***</p> | |
|--|--|

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE / FAX NO.