

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	1	Quantity	2
ABC :	P9,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>BEDSIDE TABLE & BEDSIDE CABINET TECHNICAL SPECIFICATIONS 1. Minimum dimensions: 15" x 18" x 30" 2. Hard plastic / ABS 3. With at least one (1) pull out top drawers 4. With at least two (2) layers shelves 5. With four (4) heavy duty casters Standard Requirements: Delivery 60-90 days upon receipt of Purchase Order · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable</p>			

***** Nothing Follows *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p>			
TECHNICAL SPECIFICATIONS			
Item No.:	2	Quantity	2
ABC :	P74,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>Bed, Mechanical 2 cranks with side rails and mattress</p> <p>Standard Features</p> <ul style="list-style-type: none"> · Dimensions -Bed Table -Length: at least 80 inches -Width: at least 35 inches -Height: at least 18 inches -Back Section Lifting angle: 60o-65o -Leg Section Lifting Angle: at least 30o <ul style="list-style-type: none"> - Casters with lock-5" <p>Specific Characteristics</p> <ul style="list-style-type: none"> · 2 cranks · Steel frame Epoxy Painted · Detachable ABS head and Foot Boards · Foldable cranks handle with over-loading protection system · Support Mattress Platform: ABS Molding or metal sheet · With Side Rails (at least ½ Length) Aluminum Alloy · IV socket; 4 (one on each side) · With patient name holder at foot and head <p>Accessories</p> <ul style="list-style-type: none"> · Adjustable or retractable IV pole (4 hooks) · Mattress 4"-5" thick rubberized foam with leatherette cover · Overbed table-formicatop with adjustable height · With one step foot stool <p>Standard Requirements :</p> <ul style="list-style-type: none"> · Delivery period _____ calendar days. 			

<ul style="list-style-type: none"> · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> · The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility · Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*****Nothing Follows*****</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

Specification

TELEPHONE/ FAX NO.

<p align="center">Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p>			
<p align="center">TECHNICAL SPECIFICATIONS</p>			
Item No.:	3	Quantity	6
ABC :	P33,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>BP Apparatus, Stand Type, Aneroid (with Adult & Pedia Cuff)</p> <p>Standard Accessories:</p> <ul style="list-style-type: none"> · With Stand <p>Standard Accessories:</p> <ul style="list-style-type: none"> · Bandage cuff (adult and pedia) (Velcro type) · Inflation bag (adult and pedia) (latex free) · Rubber valve with airflow control(latex free) · Preferably made in US or Germany <p>Standard Requirements :</p> <p>Delivery 60-90 days upon receipt of Purchase Order</p> <ul style="list-style-type: none"> · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p align="center">*****Nothing Follows*****</p>			

NAME OF COMPANY

ADDRESS

Technical Specifications

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

of the Philipp
ment of Health

TELEPHONE/ FAX NO.

CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	4	Quantity	6
ABC :	P102,000.00	Country of Origin:	
Brand:	Model:		
PURCHASER’S SPECIFICATION		SUPPLIER’S SPECIFICATION	
<p>DRESSING CART</p> <p>* Four pieces stainless jars with covers</p> <p>* One stainless forcep jar</p> <p>* Eight pieces reagent bottles with stopper</p> <p>* Stainless pail, bowl and tray, two drawers, shelves with railings, push handle, mounted on heavy duty casters</p> <p>* Dimension at least: W-25” L-17 ¾” H-30 ¼”</p> <p>Standard Requirements :</p> <p>Delivery 60-90 days upon receipt of Purchase Order</p> <p>· Current and Valid Certificate of Manufacturer’s compliance with ISO 13485</p> <p>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</p> <p>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</p> <p>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p>			

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	5	Quantity	1
ABC :	P25,000.00	Country of Origin:	
Brand:	Model:		
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	

EENT DIAGNOSTIC SET

Technical Specifications

Stainless Steel

Includes the following instruments:

Otoscope, with two levels of magnification

Ophthalmoscope, with different levels of magnification

Stainless steel battery handle with brightness control

Nasal speculum

Bent arm throat lamp

Illuminated tongue depressor

Post-nasal mirror 18mm diameter

Laryngeal mirror 22mm diameter

4 pcs. Rechargeable battery with charger of 220V, 60 Hz

Spare light bulb (LED/Xenon)

Standard Requirements :

- Delivery period _____ calendar days.
- Certification to provide users Manual in English Language
 - a. Operations Manual 2 copies
 - b. Service Manual 2 copies
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Training Requirement

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

After Sales Service

<ul style="list-style-type: none"> · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*** Nothing Follows ***</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p>Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p>

TECHNICAL SPECIFICATIONS			
Item No.:	6	Quantity	6
ABC :	P270,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p align="center">EMERGENCY CART</p> <p>Product Description</p> <p>Size: at least 470mm x 680mm x 980mm, 250mm cabinet height steel plate body ABS Molded base and top place Pull out writing table, push handle Four drawers with ABS Plastic basket with removable dividers At least 5mm guard rubber bumper, Four (4) caster with locking brake Accessories: 1 pc. Stainless steel trash can</p> <p>Standard Requirements :</p> <ul style="list-style-type: none"> · Delivery period _____ calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p align="center">*****Nothing Follows*****</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	7	Quantity	10
ABC :	P30,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	

EMERGENCY LIGHT (PORTABLE)

Product Description

Standard Features

- LED Lamp
- Power Supply: 220 VAC/12 VDC
- Battery: 6V/4Ah Sealed Lead-acid Free
- At least 6 operating hours (fully charged battery)

Maintenance Rechargeable Battery

Charging Indicator will be Red and Turned into Green at Full Charge

Built in Overcharge Circuit

Standard Requirements :

- Delivery period _____ calendar days.
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service

Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;">Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p>			
TECHNICAL SPECIFICATIONS			
Item No.:	8	Quantity	15
ABC :	P67,500.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>Examining Lights/Drop Lights Specifications: <ul style="list-style-type: none"> · Flexible light head · Power supply: 220 VAC, 60Hz · Type: Mobile · Bulb, 100 watts equivalent LED White light · Adjustable height: at least 400mm Material: <ul style="list-style-type: none"> · Aluminum Cast Alloy Five Stroke Base Castor with at least two (2) Castor Locks/Support Standard Requirements : <ul style="list-style-type: none"> · Delivery period _____ calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery </p>			

<p>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	9	Quantity	6
ABC :	P840,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	

High Speed Sterilizer (Table top)

Type: Water Filter

Feature: Without Bag

Application: Dry

Features:

1. The autoclave strictly meets the standard of EN13060, 93/42/CE and adopts European B standard
2. Super fast speed of the whole process, only 6 minutes
3. Five times per-vacuum preceding vacuum drying, the vacuum measurement reaches-0.098Mpa and residual humidity is less than 0.2%, suitable for miscellaneous endoscopes, instrument and longer pipes
4. 3-fixed, 1- customized program.
239Fahrenheit, 249.8 Fahrenheit, 273.2 Fahrenheit. The sterilization time and vacuum time can be adjusted by manual, ensure user obtain best sterilization effects according to private requirements
5. Microprocessor controlled. Malfunction detection system, corresponding error code will be showed on the display once there's something goes wrong
6. Having passed B/D test, the leak rate of vacuum is less than 0.13Kpa/min
7. Double safety locking device prevents door from opening while chamber is pressurized
8. An automatic safety shut off for self-protection in case of excessive temperature and pressure

Volume :10liter

Voltage Power: 220V/55Hz

Standard Requirements:

·Current and Valid Certificate of Manufacturer's compliance with ISO and CE Certificate

Warranty Certificate

·One (1) Year for craftsmanship, after inspection and acceptance

Delivery 60 days upon receipt of Purchase Order
Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

***** NOTHING FOLLOWS *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	10	Quantity	1
ABC :	P40,000.00	Country of Origin:	
Brand:	Model:		
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>Laryngoscope Set (Adult/ Pedia)</p> <p>Type: Battery Operated (Alkaline) Composition: 4 pcs – Mc Intosh blades No. 1, 2, 3, 4 2 pcs – Miller blades No. 0, 1 (6pcs Miller blades -00,0,1,2,3,4) Laryngoscope handle with LED source Features: Stainless steel construction of blades Autoclavable blades Standard accessories: Lamp case 2 pcs – LED spare lamp Rechargeble battery with charger, 220V, 60 Hz Standard Requirements :</p> <ul style="list-style-type: none"> · Delivery period _____ calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery 			

<p>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;">Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p>			
TECHNICAL SPECIFICATIONS			
Item No.:	11	Quantity	8
ABC :	P32,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Nebulizer (heavy duty)</p> <p>Technical Specifications Power Supply: 220-240 VAC , 60Hz Standard Feature: Compressor (flapper type) Standard Accessories: 2 pcs Adult mask 2 pcs Child mask 2 sets Sterilizable tubings 2 pcs. air Filter 2 pcs Mouthpiece (disposable) ABS or hard plastic body Standard Requirements : Delivery 60-90 days upon receipt of Purchase Order</p> <ul style="list-style-type: none"> · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p>			

Special Provisions

1. Non-removable DOH letters conspicuously
displayed on the front

2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	12	Quantity	4
ABC :	P4,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER’S SPECIFICATION		SUPPLIER’S SPECIFICATION	
<p>NEUROLOGICAL HAMMER</p> <p>Chrome-plated octagonal handle</p> <ul style="list-style-type: none">• Compartment for stainless steel needle and brush• Black rubber bumper• Overall length: 7” <p>Standard Requirements :</p> <ul style="list-style-type: none">• Delivery period _____ calendar days.• Current and Valid Certificate of Manufacturer’s compliance with ISO 13485• Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier• Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery• The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none">• The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility• Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none">• Bidder’s certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.			

<p>· List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions:</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*****Nothing Follows*****</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p align="center">Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON</p>			
<p align="center">TECHNICAL SPECIFICATIONS</p>			
Item No.:	13	Quantity	2
ABC :	P6,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>OXYGEN TANK CARRIER (SINGLE) Materials: Tubular Steel frame (Enamel Coated) Features · tank holder (Single Tank, 50 Lbs.) · Two heavy duty casters Standard Requirements : · Delivery period _____ calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user Special Provisions</p>			

<p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	14	Quantity	92
ABC :	P46,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Patient Chart Aluminum</p> <p>Size Long (14 x 9")</p> <p>Standard Requirements :</p> <ul style="list-style-type: none"> · Delivery period _____ calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*****Nothing Follows*****</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	15	Quantity	2
ABC :	P12,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
PATIENT CHART HOLDER (12 CAPACITY) Made of Plain sheet G-22 Chart file 12 capacity Cabinet type with Drawer & Rubber Bumper. Mounted on four wheels caster White paint or stainless Size L: 22" x H: 33.5" x W: 23" Standard Requirements : Delivery 60-90 days upon receipt of Purchase Order · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user Special Provisions			

<p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	16	Quantity	10
ABC :	P500,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER’S SPECIFICATION		SUPPLIER’S SPECIFICATION	
<p>PULSE OXIMETER</p> <p>SPO2, Pulse Rate with Warning Indicators & Pletysmography</p> <p>Audible/ Visual Alarms for High/Low SPO2 and PR, Audible/ Visual Alarms for Low-Battery</p> <p>Power Requirements: 220VAC, 60Hz,</p> <p>Table top</p> <p>Transducers for adult, neonate and pedia</p> <p>With battery back-up for at least two (2) hours</p> <p>Standard Requirements :</p> <p>Delivery 60-90 days upon receipt of Purchase Order</p> <ul style="list-style-type: none">· Current and Valid Certificate of Manufacturer’s compliance with ISO 13485· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p>			

2. Property tagging (sticker) rewritable
*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	17	Quantity	6
ABC :	P78,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER’S SPECIFICATION		SUPPLIER’S SPECIFICATION	
REFRIGERATOR SINGLE DOOR 6.5 CU FT. Type: Single Door Size: Dimension: 1234 x 525 x 560 (mm) Capacity: Capacity (Total): 149.3L Freezer: 25L Refrigerator: 119L Other: 5.3L Standard Requirements : <ul style="list-style-type: none">· Delivery period 60– 90 calendar days.· Certification to provide users Manual in English Language<ul style="list-style-type: none">a. Operations Manual 2 copiesb. Service Manual 2 copies· Current and Valid Certificate of Manufacturer’s compliance with ISO 13485· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Training Requirement <ul style="list-style-type: none">· The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility			

<ul style="list-style-type: none"> · Certificate of training must be noted/signed by the head of Facility After Sales Service · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*** Nothing Follows ***</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p>			
TECHNICAL SPECIFICATIONS			
Item No.:	18	Quantity	9
ABC :	P45,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">STETHOSCOPE ADULT</p> <p>Standard Features:</p> <ul style="list-style-type: none"> · Single-sided Chest piece (Wt: 3 - 4 ounce) · Black, Plated and Polished alloy Chest Piece · Tunable diaphragm made with epoxy or fiberglass (Dm: 1.5 - 2 inches) · Soft sealing rubberized ear tips · Rubberized tubings · Head set wide diameter aerospace alloy · Length: 25 - 30 inches · Weight: 150 - 170 gm. <p>Standard Accessories:</p> <ul style="list-style-type: none"> · 1 pair Spare Ear tips · Spare Diaphragm <p>Standard Requirements :</p> <p>Delivery 60-90 days upon receipt of Purchase Order</p> <ul style="list-style-type: none"> · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery 			

<p>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	19	Quantity	1
ABC :	P6,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>STETHOSCOPE , PEDIA & NEONATE</p> <p>Standard Features:</p> <ul style="list-style-type: none"> · Double-sided Chest piece (Wt:1 - 1.5 ounce) · Machined Stainless Steel Chest Piece · Diaphragm made with epoxy or fiberglass (Dm: 2.5 - 3.3 cm) · Soft sealing rubberized ear tips · Rubberized tubings · Head set wide diameter aerospace alloy · Length: 70 - 75 cm · Weight: 100 - 110 gm. <p>Standard Accessories:</p> <ul style="list-style-type: none"> · 1 pair Spare Ear tips · Spare Diaphragm <p>Standard Requirements :</p> <p>Delivery 60-90 days upon receipt of Purchase Order</p> <ul style="list-style-type: none"> · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p>			

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p>			
TECHNICAL SPECIFICATIONS			
Item No.:	20	Quantity	5
ABC :	P82,500.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">WEIGHING SCALE WITH MEASURING STICK, ADULT</p> <p> Mechanical weighing scale with height rod 220kgs maximum weighing capacity Non-slip steel platform Adjustable/telescopic height rod with 1mm graduation Engraved numbers on eye level scale beam and height rod Large low rise platform (at least 13 inches length x 13 inches x 3 inches high Weight increments/graduation of 50grams Height rod of aluminum material Powder coated body Standard Requirements : Delivery 60-90 days upon receipt of Purchase Order · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Warranty Certificate </p>			

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*****Nothing Follows*****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	21	Quantity	16
ABC :	P128,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Wheel Chair (IV STAND)</p> <p>Standard Features Heavy duty, locally made, stainless steel or chrome plated, durable Foldable Leatherette or equivalent seat and back rest foot rest wheel lock handle Standard Requirements : Delivery 60-90 days upon receipt of Purchase Order · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user Special Provisions 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable *****Nothing Follows*****</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON			
TECHNICAL SPECIFICATIONS			
Item No.:	22	Quantity	14
ABC :	P770,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER’S SPECIFICATION		SUPPLIER’S SPECIFICATION	
<div>IV INFUSION PUMP</div> <div>Power Specifications</div> <div><div>· Power Supply: 220VAC, 60Hz</div><div>· Battery: (rechargeable)</div><div>· Battery Life: at least 3 hours at 30 ml/h</div></div> <div>General Specifications</div> <div><div>· Automatic Infusion</div><div>· Flow rate: capable of 1-1100 ml/h</div><div>· Visible/audible alarms Audio/Visual Alarm</div><div>a. occlusion</div><div>b. Air in line</div><div>c. Door open</div><div>d. End program</div><div>e. Low pressure</div><div>· With optional data interface port of infusion supervision system</div><div>· May or maybe upgradable to PC connectivity for central monitoring</div><div>· With drug delivery chart</div></div> <div>Standard Requirements :</div> <div>Delivery 60-90 days upon receipt of Purchase Order</div> <div><div>· Current and Valid Certificate of Manufacturer’s compliance with ISO 13485</div><div>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</div><div>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</div></div>			

<p>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</p> <p>Warranty Certificate</p> <p>Minimum of one (1) year for Parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <p>1. Non-removable DOH letters conspicuously displayed on the front</p> <p>2. Property tagging (sticker) rewritable</p> <p>*****Nothing Follows*****</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p>			
TECHNICAL SPECIFICATIONS			
Item No.:	23	Quantity	9
ABC :	P1,800,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">SUCTION MACHINE 10L</p> <p>Max vacuum flow up to 720-740mmHg on average</p> <p>Features one piece molded outer body for superior resistance to rust</p> <p>Stainless steel Tray on top convenient for tools treatment</p> <p>Comes with 2 glass collection jars each 3000ml with cover CE mark complies with EEC/EMC of EN60601-1-2 Requirement Ideal for suction environment in operation theater room</p> <p>Standard accessories Silicon Tube, Curette Tube set, Filter included</p> <p>Standard Requirements :</p> <ul style="list-style-type: none"> · Delivery period 60– 90 calendar days. · Certification to provide users Manual in English Language a. Operations Manual 2 copies b. Service Manual 2 copies · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient 			

<p>and respective regional offices prior to the delivery</p> <ul style="list-style-type: none"> · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> · The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility · Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*** Nothing Follows ***</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

