

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>1</b>	<b>Quantity</b>	<b>2</b>
<b>ABC :</b>	<b>P1,000,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>HEARING TEST MACHINE</b></p> <p><b>Specifications:</b>            Measurement techniques TEOAE  <b>TEOAE</b>            Evaluation method: Nose weighted averaging, counting of significant signal peaks            Stimulus: Non-linear click sequence            Stimulus level: 70-84 dB SPL (45-60 dB HL), self calibration depending on ear canal volume            Click rate: Approx. 60 Hz            Frequency rate: 1.5 to 4.5 kHz            Display: Statistical waveform, measurement progress, TEOAE detection level, noise level  <b>Display</b>            Type: Color, TFT, touch screen with adjustable LED backlight            Dimensions: Manufacturer Default            Resolution: 240 x 320 pixels            Keystroke durability: min. 1 million repetitive strokes per screen point  <b>Keypad:</b> Resistive Touch Screen (can be used with gloves)  <b>Memory:</b> Patient memory capacity: Max. 250 patients / Min. 500 tests  <b>Connectors:</b> OAE probe connector: 14 pin ODU Medisnap  <b>Real time clock:</b>            - Integrated real-time clock for time-stamping of measurements.            - The clock is automatically synchronized with PC clock when docked.  <b>Backup:</b> Min. 5 days, when battery is removed from unit  <b>Data interfaces</b>            PC: IR data transmission to Docking Station -            USB interface from Docking Station to PC  <b>Transport and storage environment</b></p>			

Temperature range: -20 - +60°C (-4 - 140 °F)  
Humidity range: 20-80% rel., non-condensing  
Air pressure 500 hPa to 1060 hPa

**Operating environment:**  
Temperature range: 10 - 40 °C (50 - 104°F)  
Humidity range: 30-80% rel., non-condensing  
Air pressure 600 hPa to 1060 hPa GN  
Otometrics

**Standards**  
Otoacoustics emissions: EN 60645-6, Type 2  
Auditory evoked potentials: EN 60645-7, Type 2

**Patient Safety:**  
•EN 60601-1, Internally Powered, Type BF, IPXO  
•UL 60601-1  
• IEC 60601-2-26  
• IEC 60601-2-40 EMC: EN 60601-1-2

**Power supply and battery**  
Battery type: Rechargeable Li-ion 3.7 V/1800 mAh (6.7 Wh), fully charged  
Actual use can influence the battery life time  
Estimated battery life: 8 hours of continuous use (based on a typical use scenario)  
Battery level indicator: 5 step battery level indicator  
Charge time in docking station: 80% charged 4 ½ hours. Fully charged 6 hours.

**PC interface type:**  
Interface type: USB 2.0, Full-speed  
USB Power: Uses <100 mA of current from the USB interface

**Printer interface**  
Interface type: RS232  
Connector type: 6-pol Mini Din

**DC power input**  
Input voltage: 5 V DC ±5%  
Max. power consumption when AccuScreen is docked: 5VA (5V, 1.0 A)  
Max. power consumption when AccuScreen is not docked: 0.25 VA (5V, 50 mA)

**Power adapter:**  
Input voltage/range: 100-240 V AC, 50-60 Hz  
Output voltage: 5.0 V DC/min. 1.0A  
Mains plug types: US, UK, Europe and Australia

**Probe:** Flexible, shielded cable, approx. length: 120 cm (approx. 55 inches)

**Dimensions:** Probe body: 20 mm Ø x 23 x 11 mm (0.8" Ø x 0.9" x 0.43")  
Probe tip: 3.3 mm Ø x 10 mm (0.13" Ø x 0.4")

**Weight:** Probe incl. probe tip: Approx. 4.5 grams

**Eartips**

Standard (cylindric): 4 sizes (3.7 - 5 mm)

Tree tip: 1 size (4 - 7 mm)

Foam tip: 1 size (13 mm) ABR electrode cable

**Standard accessories**

Docking Station

Power adaptor and USB cable

Starter kit

OAE probe (DPOAE/TEOAE or TEOAE only)

Battery

ABR electrode cable (ABR version only)

ABR Tester (ABR version only)

**Standard Requirements:**

- Delivery period: 100 calendar days.
- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

**Training Requirement**

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

**After Sales Service**

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

**Warranty Certificate**

Warranty services for 1 year. Warranty includes labor and parts, free of charge. However, this excludes damages due to negligent acts of the user who shall be held liable for the cost of

repair and spare parts. Only appointed service engineers shall be allowed to service the machine.

**Special Provisions:**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

**\*\*\* Nothing Follows \*\*\***

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>2</b>	<b>Quantity</b>	<b>10</b>
<b>ABC :</b>	<b>P550,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>ER Mobile bed</b></p> <p><b>Product Description</b>  Hard deck for rigorous medical procedures  Adjustable backrest  Single-hook IV pole  Removable side rails  Anti-microbial mattress  Set of 4 locking casters  Maximum of 500 lbs load capacity  Sleeping surface: Manufacturer's Default  Height: 19" - 32" (adjustable)  Stored: Manufacturer's Default  Weight: Manufacturer's Default</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period 100 calendar days.</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b>  Minimum of one (1) year for Parts &amp; service  Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p>			

- |  |  |
|--|--|
| <ol style="list-style-type: none"><li>1. Non-removable DOH letters conspicuously displayed on the front</li><li>2. Property tagging (sticker) rewritable</li></ol> <p><b>*** Nothing Follows ***</b></p> |  |
|--|--|

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>3</b>	<b>Quantity</b>	<b>15</b>
<b>ABC :</b>	<b>P525,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>EXAMINING TABLE WITH STIRRUPS</b></p> <p><b>TECHNICAL SPECIFICATIONS</b></p> <p><b>Standard Features:</b></p> <ol style="list-style-type: none"> <li>1. steel top 3 section adjustable</li> <li>2. with stainless steel heel detachable stirrups</li> <li>3. with mattress &amp; leatherette cover at least 4"</li> <li>4. with stainless steel drain pan</li> <li>5. with rubber covered footstep</li> </ol> <p>Table Top Dimensions: Manufacturer's Default</p> <ol style="list-style-type: none"> <li>1. length: Manufacturer's Default</li> <li>2. width: Manufacturer's Default</li> <li>3. height: Manufacturer's Default</li> </ol> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days</li> <li>· Certification to provide user's Manual in English Language               <ol style="list-style-type: none"> <li>a. Operations Manual 2 copies</li> <li>b. Service Manual 2 copies</li> </ol> </li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of three (3) years for both parts &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p>			

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
  2. Property tagging (sticker) rewritable
- \*\*\* Nothing Follows \*\*\***

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.



# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>4</b>	<b>Quantity</b>	<b>6</b>
<b>ABC :</b>	<b>P270,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>	<b>Model:</b>		
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>EMERGENCY CART</b></p> <p><b>Product Description</b>            Size: at least 470mm x 680mm x 980mm,            250mm cabinet height steel plate body            ABS Molded base and top place            Pull out writing table, push handle Four drawers            with ABS Plastic basket            with removable dividers            At least 5mm guard rubber bumper,            Four (4) caster with locking brake  <b>Accessories:</b>            1 pc. Stainless steel trash can  <b>Standard Requirements:</b>            · Delivery period: 100 calendar days.            · Current and Valid Certificate of Manufacturer's            compliance with ISO 13485            · Certificate of Commitment from the legal            manufacturer of the unit that they will honor the            warranty in case of change of distributor/supplier            · Transportation to site, delivery, installation and            testing expenses on the site (health facility) for            the account of the supplier and notification of the            supplier to the recipient and respective regional            offices prior to the delivery            · The manufacturers certificate that the brand            must be in the local market for at least five (5)            years and 10 years in the international market  <b>Warranty Certificate</b>            Minimum of one (1) year for craftsmanship &amp;            service Warranty Certificate upon delivery,            inspection and acceptance and after the provision            of the required training of the end-user  <b>Special Provisions</b></p>			

- |  |  |
|--|--|
| <ol style="list-style-type: none"><li>1. Non-removable DOH letters conspicuously displayed on the front</li><li>2. Property tagging (sticker) rewritable</li></ol> <p><b>*** Nothing Follows ***</b></p> |  |
|--|--|

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>5</b>	<b>Quantity</b>	<b>13</b>
<b>ABC :</b>	<b>P195,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>INSTRUMENT TABLE</b></p> <p><b>Type:</b> Mobile</p> <p><b>Material (Make):</b> Stainless Steel</p> <p><b>Table top:</b></p> <p>Size (minimum)</p> <p><b>Length – at least 80 cm</b></p> <p><b>Width – at least 50 cm</b></p> <p><b>Height – at least 90 cm</b></p> <p><b>Features:</b></p> <p>With 4 caster</p> <p>Stainless Steel Tray</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days.</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of one (1) year for craftsmanship &amp; service</p> <p>Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p style="text-align: center;">*** <b>Nothing Follows</b> ***</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>6</b>	<b>Quantity</b>	<b>13</b>
<b>ABC :</b>	<b>P195,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>GANG CHAIR</b></p> <p><b>Standard Features</b></p> <ul style="list-style-type: none"> <li>· Material: Stainless Steel</li> <li>· Dimension: 60 cm</li> </ul> <p>3 seater capacity, double the quantity</p> <ul style="list-style-type: none"> <li>· With backrest</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO and/or CE Certificate</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Training Requirement</b></p> <ul style="list-style-type: none"> <li>· The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility</li> <li>· Certificate of training must be noted/signed by the head of Facility</li> </ul> <p><b>After Sales Service</b></p> <ul style="list-style-type: none"> <li>· Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.</li> <li>· List of service centers in major cities where equipment is installed and supplies / reagent</li> </ul>			

outlet with contact landline numbers should be available

**Warranty Certificate**

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

**\*\*\* Nothing Follows \*\*\***

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>7</b>	<b>Quantity</b>	<b>7</b>
<b>ABC :</b>	<b>P175,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>EENT DIAGNOSTIC SET</b></p> <p><b>Technical Specifications</b></p> <p>Stainless Steel</p> <p><b>Includes the following instruments:</b></p> <p>Otoscope, with two levels of magnification</p> <p>Ophthalmoscope, with different levels of magnification</p> <p>Stainless steel battery handles with brightness control</p> <p>Nasal speculum</p> <p>Bent arm throat lamp</p> <p>Illuminated tongue depressor</p> <p>Post-nasal mirror 18mm diameter</p> <p>Laryngeal mirror 22mm diameter</p> <p>4 pcs. Rechargeable battery with charger of 220V, 60 Hz</p> <p>Spare light bulb (LED/Xenon)</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days</li> <li>· Certification to provide user's Manual in <b>English Language</b> <ul style="list-style-type: none"> <li>a. Operations Manual 2 copies</li> <li>b. Service Manual 2 copies</li> </ul> </li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Training Requirement</b></p>			

<ul style="list-style-type: none"> <li>· The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility</li> <li>· Certificate of training must be noted/signed by the head of Facility</li> </ul> <p><b>After Sales Service</b></p> <ul style="list-style-type: none"> <li>· Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.</li> <li>· List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available</li> </ul> <p><b>Warranty Certificate</b> Minimum of three (3) years for both parts &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p><b>*** Nothing Follows ***</b></p>	
--	--

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p align="center"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<p align="center"><b>TECHNICAL SPECIFICATIONS</b></p>			
<b>Item No.:</b>	<b>8</b>	<b>Quantity</b>	<b>10</b>
<b>ABC :</b>	<b>P170,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p align="center"><b>DRESSING CART</b></p> <ul style="list-style-type: none"> <li>* Four pieces stainless jars with covers</li> <li>* One stainless forcep jar</li> <li>* Eight pieces reagent bottles with stopper</li> <li>* Stainless pail, bowl and tray, two drawers, shelves with railings, push handle, mounted on heavy duty casters</li> <li>* Dimension at least: W-25" L-17 ¾" H-30 ¼"</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p align="center"><b>*** Nothing Follows ***</b></p>			

NAME OF COMPANY

ADDRESS



# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>9</b>	<b>Quantity</b>	<b>3</b>
<b>ABC :</b>	<b>P165,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>Wheeled stretcher, heavy duty, stainless steel with siderails, IV pole and oxygen tank carrier</b></p> <p><b>Standard Features</b></p> <p><b>Technical Specifications:</b></p> <ul style="list-style-type: none"> <li>· Minimum dimensions:</li> <li>Width: 740mm</li> <li>Length: 1900mm</li> <li>Height: 750mm</li> <li>· Mounted on four (4) swivel castors with brakes</li> <li>· Filled mattress pad with leatherette cover 4” thick of better</li> <li>· I.V. pole holder</li> <li>· Push bar</li> <li>· Rubber bumper</li> <li>· Adjustable height</li> <li>· Manual crank</li> <li>· Mini O2 tank</li> <li>· Collapsible side rails</li> </ul> <p>Standard accessories (2 pcs Restrain strap)</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p>			

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

\*\*\* **Nothing Follows** \*\*\*

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

Republic of the Philippines  
Department of Health  
CENTER FOR HEALTH DEVELOPMENT CALABARZON

## TECHNICAL SPECIFICATIONS

<b>Item No.:</b>	<b>10</b>	<b>Quantity</b>	<b>33</b>
<b>ABC :</b>	<b>P148,500.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>Examining Lights/Drop Lights</b></p> <p><b>Specifications:</b></p> <ul style="list-style-type: none"> <li>· Flexible light head</li> <li>· Power supply: 220 VAC, 60Hz</li> <li>· Type: Mobile</li> <li>· Bulb, 100 watts equivalent LED White light</li> <li>· Adjustable height: at least 400mm</li> </ul> <p><b>Material:</b></p> <ul style="list-style-type: none"> <li>· Aluminum Cast Alloy Five Stroke Base Castor with at least two (2) Castor Locks/Support</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days .</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p style="text-align: center;"><b>*** Nothing Follows ***</b></p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>11</b>	<b>Quantity</b>	<b>2</b>
<b>ABC :</b>	<b>P140,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p><b>ECG MACHINE WITH ANALYZER</b></p> <p><b>Standard Features</b></p> <ul style="list-style-type: none"> <li>· Lead: 12 Leads</li> <li>· Color Display</li> <li>· Built-in memory of at least 100 patients in auto mode</li> <li>· Patient data can be stored on USB flash drive</li> <li>· Power Supply: 220V-60Hz with Battery Pack</li> <li>· Print name/ID label, HR, axis, interval, voltage measurement, averages and interpretation</li> <li>· Minimum printing width of 90 mm</li> <li>· Defibrillator and Electro Magnetic Interference protected</li> <li>· HL7 compatible for data transfer</li> <li>· Shock proof</li> </ul> <p><b>Accessories</b></p> <ul style="list-style-type: none"> <li>· Patient Cable</li> <li>· Reusable Patient Electrodes (Limb and Chest)</li> <li>· Power Cord</li> <li>· Ground Wire</li> <li>· Carrying Case</li> <li>· Include Cart/Trolley with drawer and 2 or 4 lockable casters</li> </ul> <p><b>Consumables</b></p> <ul style="list-style-type: none"> <li>· ECG recording paper, 1 roll</li> <li>· ECG Gel</li> </ul> <p><b>Parameters</b></p> <ul style="list-style-type: none"> <li>· HHHR, P-R interval, P duration, QRS duration, T duration, Q-T interval, Q-Tc, P axis, QRS, T axis, R (V5), S (V1), R (V5) + S (V1)</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days .</li> <li>· Certification to provide user's Manual in</li> </ul> <p><b>English Language</b></p> <ol style="list-style-type: none"> <li>Operations Manual 2 copies</li> <li>Service Manual 2 copies</li> </ol>			

- Current and Valid Certificate of Manufacturer's compliance with ISO 13485
- Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier
- Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery
- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

#### **Training Requirement**

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

#### **After Sales Service**

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

#### **Warranty Certificate**

Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

#### **Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

\*\*\* **Nothing Follows** \*\*\*

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p align="center"><b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b></p>			
<p align="center"><b>TECHNICAL SPECIFICATIONS</b></p>			
<b>Item No.:</b>	<b>12</b>	<b>Quantity</b>	<b>17</b>
<b>ABC :</b>	<b>P136,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p align="center"><b>Wheel Chair with I.V. Stand</b></p> <p><b>Standard Features</b>  Heavy duty, locally made, stainless steel or chrome plated, durable  Foldable  Leatherette or equivalent seat and back rest  foot rest  wheel lock  handle</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period:100 calendar days .</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b>  Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p align="center">*** Nothing Follows ***</p>			

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p align="center"><b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b></p>			
<p align="center"><b>TECHNICAL SPECIFICATIONS</b></p>			
<b>Item No.:</b>	<b>13</b>	<b>Quantity</b>	<b>24</b>
<b>ABC :</b>	<b>P96,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p align="center"><b>Nebulizer</b></p> <p><b>Technical Specifications</b>  Power Supply:220-240 VAC , 60Hz  <b>Standard Feature:</b>  Compressor (flapper type)  <b>Standard Accessories:</b>  2 pcs Adult mask  2 pcs Child mask  2 sets Sterilizable tubings  2 pcs. air Filter  2 pcs Mouthpiece (disposable)  ABS or hard plastic body  <b>Standard Requirements:</b>  · Delivery period:100 calendar days .  · Current and Valid Certificate of Manufacturer's compliance with ISO 13485  · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier  · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery  · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market  <b>Warranty Certificate</b>  Minimum of one (1) year for craftsmanship &amp; service  Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user  <b>Special Provisions</b>  1. Non-removable DOH letters conspicuously displayed on the front  2. Property tagging (sticker) rewritable  *** <b>Nothing Follows</b> ***</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"><b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b></p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>14</b>	<b>Quantity</b>	<b>4</b>
<b>ABC :</b>	<b>P80,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p><b>EXAMINING TABLE WITH NO STIRRUPS</b>  <b>TECHNICAL SPECIFICATIONS</b>  <b>Standard Features:</b>            1. steel top 3 section adjustable            2. with mattress &amp; leatherette cover at least 4"            3. with stainless steel drain pan            4. with rubber covered footstep  <b>Table Top Dimensions:</b>            1. length: at least 1700mm            2. width: at least 590mm            3. height: at least 650mm  <b>Standard Requirements:</b>  <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days.</li> <li>· Certification to provide user's Manual in English Language</li> <li>a. Operations Manual 2 copies</li> <li>b. Service Manual 2 copies</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <b>Training Requirement</b>  <ul style="list-style-type: none"> <li>· The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility</li> <li>· Certificate of training must be noted/signed by the head of Facility</li> </ul> </p>			



**After Sales Service**

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.

- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

**Warranty Certificate**

Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

\*\*\*\*NOTHING FOLLOWS\*\*\*\*

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p align="center"><b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b></p>			
<p align="center"><b>TECHNICAL SPECIFICATIONS</b></p>			
<b>Item No.:</b>	<b>15</b>	<b>Quantity</b>	<b>4</b>
<b>ABC :</b>	<b>P80,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
<b>PURCHASER'S SPECIFICATION</b>		<b>SUPPLIER'S SPECIFICATION</b>	
<p align="center"><b>Suction Machine 2L capacity</b></p> <p><b>Mobile</b>  Power Supply : 220 VAC, 6Hz  Capacity : 2 liters or better</p> <p><b>Features</b>  Vacuum Pressure: 700mmHg (min)  Suction jars- 2 pieces  With 2 m long suction hose with nipple  Provision of Gauge and regulator</p> <p><b>Standard Accessory</b>  Plastic Tubing</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days.</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b>  Minimum of one (1) year for craftsmanship &amp; service  Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p align="center"><b>*** Nothing Follows ***</b></p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>16</b>	<b>Quantity</b>	<b>1</b>
<b>ABC :</b>	<b>P50,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>PULSE OXIMETER (Table Top)</b></p> <p>SPO2, Pulse Rate with Warning Indicators &amp; Pletysmography</p> <p>Audible/ Visual Alarms for High/Low SPO2 and PR, Audible/ Visual Alarms for Low-Battery</p> <p>Power Requirements: 220VAC, 60Hz,</p> <p>Table top</p> <p>Transducers for adult, neonate and pedia</p> <p>With battery back-up for at least two (2) hours</p> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days.</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p style="text-align: center;">*** Nothing Follows ***</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

# Technical Specifications

<p align="center"><b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b></p>			
<p align="center"><b>TECHNICAL SPECIFICATIONS</b></p>			
<b>Item No.:</b>	<b>17</b>	<b>Quantity</b>	<b>2</b>
<b>ABC :</b>	<b>P40,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p><b>Tactical Traction Splint (TTS) NAR</b>          -Lightweight and portable          -Clearly marked elastic straps for the proper fit to the thigh, knee and shin          Dimensions:          Packed: L 11 in. x W 5.5 in. x D 3 in.          Weight: 1 lb 5 oz  <b>Standard Requirements:</b>          · Delivery period: 100 calendar days.          · Current and Valid Certificate of Manufacturer's compliance with ISO 13485          · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier          · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery          · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market  <b>Warranty Certificate</b>          Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user  <b>Special Provisions</b>          1. Non-removable DOH letters conspicuously displayed on the front          2. Property tagging (sticker) rewritable          *** <b>Nothing Follows</b> ***</p>			

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
 AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>18</b>	<b>Quantity</b>	<b>1</b>
<b>ABC :</b>	<b>P10,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>Spine Board complete set w/straps</b></p> <p><b>Dimensions and Weight: Manufacturer Default</b></p> <p><b>Thickness:</b> • 2.3 inches</p> <p><b>Weighs: Manufacturer Default</b></p> <p><b>Operating Temperature:</b></p> <p>• -18°C to 43°C (0°F to 110°F)</p> <p><b>Storage Temperature:</b></p> <p>• -18°C to 43°C (0°F to 110°F)</p> <p><b>Standard Features:</b></p> <ul style="list-style-type: none"> <li>• 14 handholds</li> <li>• Can be used for water rescue, in the mountains, at home or on the road</li> <li>• Extra-large handholds</li> <li>• Raised handholds</li> <li>• Pins are molded into the board</li> <li>• Includes strap holes for pediatric patients</li> <li>• Minimum storage space for 2 boards of approximated 4 in.</li> <li>• Radiolucent for clear view of the spine, MRI and CT scan compatible</li> <li>• Durable molded-in customized labeling available</li> <li>• 12 Pins</li> <li>• Carbon Fiber Tubes</li> <li>• Latex-free</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>• Delivery period 100 calendar days.</li> <li>• Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>• Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>• Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the</li> </ul>			

supplier to the recipient and respective regional offices prior to the delivery

· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

**Warranty Certificate**

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

**\*\*\* Nothing Follows \*\*\***

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>19</b>	<b>Quantity</b>	<b>2</b>
<b>ABC :</b>	<b>P6,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>HEAD IMMOBILIZER, ADULT</b></p> <p> Instrument classification: Class I  Material: High density plastic  Size: Manufacturer's Default  Certification: CE FDA ISO:9001/ISO:13485  Color: Red, blue  Net Weight: 1kg  Feature: used with scoop stretcher or spine board; can be adjusted  Application: Emergency rescue for adult  Warranty: 2 Years  Type: First-Aid Devices  Foamed plastics with Polyethylene Coated type:  Standard head immobilizer material:  high density foam material, water repellent structure:  two pillows, 2 removable straps, 1 basic pillow, 2 straps  color:  Red or blue or yellow for general  package:  Single package  dimension:  43.5 x 27x 14cm  weight:  2KG  <b>Standard Requirements:</b>  · Delivery period 100 calendar days.  · Current and Valid Certificate of Manufacturer's compliance with ISO 13485  · Certificate of Commitment from the legal manufacturer of the unit that they will <b>honor the</b> warranty in case of change of distributor/supplier  · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the </p>			

supplier to the recipient and respective regional offices prior to the delivery

- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

**Training Requirement**

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

**After Sales Service**

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

**Warranty Certificate**

Warranty services for 1 year. Warranty includes labor and parts, free of charge. However, this excludes damages due to negligent acts of the user who shall be held liable for the cost of repair and spare parts. Only appointed service engineers shall be allowed to service the machine.

**Special Provisions:**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

**\*\*\* Nothing Follows \*\*\***

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.



# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>20</b>	<b>Quantity</b>	<b>1</b>
<b>ABC :</b>	<b>P400,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>AUTOMATIC FILM PROCESSOR</b></p> <p> <b>Power Supply: 220V</b>  <b>Console Advance</b>  Imaging Plate ST-VI:  <b>35 x 43 cm (14" x 17") ,x 35 x 35cm</b>  <b>(14"x14"), 10"x12",8"x10"</b>  <b>Min. 49 sec</b>  <b>Up to 73 Ips/hr</b>  <b>IP Cassete Type CC:</b>  <b>10 pixels / mm, 5 pixels /mm</b>  <b>Min. 33 sec.</b>  <b>Approx 165 sec. (Approx. 155 sec) * in case of</b>  <b>35 x 43cm (14" x 17")</b>  <b>1</b>  10 Base T/100 Base TX  560 x 540 x 392 mm (22" x 21" x 15")  39 kg (86 lbs)  Single phase 50-60 hz  AC 120-240 V = 10% 1.9A (max)  <b>Operating Conditions:</b>  * Temperature: 15-30 C  * Humidity : 15-80% RH (No dew condensation)  * Atmosphere pressure: 750-1060hPa  <b>Standard Requirements :</b>  · Delivery period 100 calendar days.  · Certification to provide user's Manual in English Language  a. Operations Manual 2 copies  b. Service Manual 2 copies  · Current and Valid Certificate of Manufacturer's compliance with ISO 13485  · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier  · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the </p>			

supplier to the recipient and respective regional offices prior to the delivery

- The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

**Training Requirement**

- The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility
- Certificate of training must be noted/signed by the head of Facility

**After Sales Service**

- Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.
- List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

**Warranty Certificate**

Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

**Special Provisions**

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

\*\*\* **Nothing Follows** \*\*\*

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

# Technical Specifications

<p style="text-align: center;"> <b>Republic of the Philippines</b>  <b>Department of Health</b>  <b>CENTER FOR HEALTH DEVELOPMENT CALABARZON</b> </p>			
<b>TECHNICAL SPECIFICATIONS</b>			
<b>Item No.:</b>	<b>21</b>	<b>Quantity</b>	<b>6</b>
<b>ABC :</b>	<b>P90,000.00</b>	<b>Country of Origin:</b>	
<b>Brand:</b>		<b>Model:</b>	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;"><b>NEGATOSCOPE</b></p> <p><b>Standard Features:</b></p> <ul style="list-style-type: none"> <li>- DOUBLE FRAME)</li> <li>- Steel body; min. measurement 127mmhtx774mmwide wall mounted, sufficient and even light distribution, durable sliding clip (preferably magic clip) supplied with film hanger, X-RAY illuminator, double screen w/ stand and wheel painted</li> </ul> <p><b>Standard Requirements:</b></p> <ul style="list-style-type: none"> <li>· Delivery period: 100 calendar days.</li> <li>· Current and Valid Certificate of Manufacturer's compliance with ISO 13485</li> <li>· Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier</li> <li>· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery</li> <li>· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market</li> </ul> <p><b>Warranty Certificate</b></p> <p>Minimum of one (1) year for craftsmanship &amp; service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p><b>Special Provisions</b></p> <ol style="list-style-type: none"> <li>1. Non-removable DOH letters conspicuously displayed on the front</li> <li>2. Property tagging (sticker) rewritable</li> </ol> <p style="text-align: center;">*** <b>Nothing Follows</b> ***</p>			

---

NAME OF COMPANY

---

ADDRESS

---

SIGNATURE OVER PRINTED NAME OF  
AUTHORIZED REPRESENTATIVE

---

TELEPHONE/ FAX NO.

