

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	1	Quantity	10
ABC :	P550,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">ER Mobile bed</p> <p>Product Description Hard deck for rigorous medical procedures Adjustable backrest Single-hook IV pole Removable side rails Anti-microbial mattress Set of 4 locking casters Maximum of 500 lbs load capacity Sleeping surface: Manufacturer's Default Height: 19" - 32" (adjustable) Stored: Manufacturer's Default Weight: Manufacturer's Default</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for Parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p>			

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
 2. Property tagging (sticker) rewritable
- *** Nothing Follows *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	2	Quantity	6
ABC :	P270,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">EMERGENCY CART</p> <p>Product Description Size: at least 470mm x 680mm x 980mm, 250mm cabinet height steel plate body ABS Molded base and top place Pull out writing table, push handle Four drawers with ABS Plastic basket with removable dividers At least 5mm guard rubber bumper, Four (4) caster with locking brake</p> <p>Accessories: 1 pc. Stainless steel trash can</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period: 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p>			

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
 2. Property tagging (sticker) rewritable
- *** Nothing Follows *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	3	Quantity	3
ABC :	P165,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Wheeled stretcher, heavy duty, stainless steel with siderails, IV pole and oxygen tank carrier</p> <p>Standard Features</p> <p>Technical Specifications:</p> <ul style="list-style-type: none"> · Minimum dimensions: Width: 740mm Length: 1900mm Height: 750mm · Mounted on four (4) swivel castors with brakes · Filled mattress pad with leatherette cover 4” thick of better · I.V. pole holder · Push bar · Rubber bumper · Adjustable height · Manual crank · Mini O2 tank · Collapsible side rails <p>Standard accessories (2 pcs Restrain strap)</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period: 100 calendar days · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery 			

· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*** **Nothing Follows** ***

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	4	Quantity	2
ABC :	P140,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>ECG MACHINE WITH ANALYZER</p> <p>Standard Features</p> <ul style="list-style-type: none"> · Lead: 12 Leads · Color Display · Built-in memory of at least 100 patients in auto mode · Patient data can be stored on USB flash drive · Power Supply: 220V-60Hz with Battery Pack · Print name/ID label, HR, axis, interval, voltage measurement, averages and interpretation · Minimum printing width of 90 mm · Defibrillator and Electro Magnetic Interference protected · HL7 compatible for data transfer · Shock proof <p>Accessories</p> <ul style="list-style-type: none"> · Patient Cable · Reusable Patient Electrodes (Limb and Chest) · Power Cord · Ground Wire · Carrying Case · Include Cart/Trolley with drawer and 2 or 4 lockable casters <p>Consumables</p> <ul style="list-style-type: none"> · ECG recording paper, 1 roll · ECG Gel <p>Parameters</p> <ul style="list-style-type: none"> · HHHR, P-R interval, P duration, QRS duration, T duration, Q-T interval, Q-Tc, P axis, QRS, T axis, R (V5), S (V1), R (V5) + S (V1) <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period:100 calendar days . 			

<ul style="list-style-type: none"> · Certification to provide user's Manual in English Language <ul style="list-style-type: none"> a. Operations Manual 2 copies b. Service Manual 2 copies · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> · The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility · Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate</p> <p>Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*** Nothing Follows ***</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	5	Quantity	17
ABC :	P136,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Wheel Chair with I.V. Stand</p> <p>Standard Features Heavy duty, locally made, stainless steel or chrome plated, durable Foldable Leatherette or equivalent seat and back rest foot rest wheel lock handle</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period:100 calendar days . · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*** Nothing Follows ***</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	6	Quantity	4
ABC :	P80,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p>Suction Machine 2L capacity Mobile Power Supply : 220 VAC, 6Hz Capacity : 2 liters or better Features Vacuum Pressure: 700mmHg (min) Suction jars- 2 pieces With 2 m long suction hose with nipple Provision of Gauge and regulator Standard Accessory Plastic Tubing Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period: 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at 			

least five (5) years and 10 years in the international market

Warranty Certificate

Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user

Special Provisions

1. Non-removable DOH letters conspicuously displayed on the front
2. Property tagging (sticker) rewritable

***** Nothing Follows *****

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	7	Quantity	2
ABC :	P40,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Tactical Traction Splint (TTS) NAR</p> <p>-Lightweight and portable</p> <p>-Clearly marked elastic straps for the proper fit to the thigh, knee and shin</p> <p>Dimensions:</p> <p>Packed: L 11 in. x W 5.5 in. x D 3 in.</p> <p>Weight: 1 lb 5 oz</p> <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period: 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate</p> <p>Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p style="text-align: center;">*** Nothing Follows ***</p>			

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	8	Quantity	1
ABC :	P10,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">Spine Board complete set w/straps</p> <p>Dimensions and Weight: Manufacturer Default</p> <p>Thickness: 2.3 inches</p> <p>Weighs: Manufacturer Default</p> <p>Operating Temperature:</p> <ul style="list-style-type: none"> · -18°C to 43°C (0°F to 110°F) <p>Storage Temperature:</p> <ul style="list-style-type: none"> · -18°C to 43°C (0°F to 110°F) <p>Standard Features:</p> <ul style="list-style-type: none"> · 14 handholds · Can be used for water rescue, in the mountains, at home or on the road · Extra-large handholds · Raised handholds · Pins are molded into the board · Includes strap holes for pediatric patients · Minimum storage space for 2 boards of approximated 4 in. · Radiolucent for clear view of the spine, MRI and CT scan compatible · Durable molded-in customized labeling available · 12 Pins · Carbon Fiber Tubes · Latex-free <p>Standard Requirements:</p> <ul style="list-style-type: none"> · Delivery period 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier 			

<ul style="list-style-type: none"> · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Warranty Certificate Minimum of one (1) year for craftsmanship & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*** Nothing Follows ***</p>	
---	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

Republic of the Philippines
Department of Health
CENTER FOR HEALTH DEVELOPMENT CALABARZON

TECHNICAL SPECIFICATIONS

Item No.:	9	Quantity	2
ABC :	P6,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
HEAD IMMOBILIZER, ADULT Instrument classification: Class I Material: High density plastic Size: Manufacturer's Default Certification: CE FDA ISO:9001/ISO:13485 Color: Red, blue Net Weight: 1kg Feature: used with scoop stretcher or spine board; can be adjusted Application: Emergency rescue for adult Warranty: 2 Years Type: First-Aid Devices Foamed plastics with Polyethylene Coated type: Standard head immobilizer material: high density foam material, water repellent structure: two pillows, 2 removable straps, 1 basic pillow, 2 straps color: Red or blue or yellow for general package: Single package dimension: 43.5 x 27x 14cm weight: 2KG Standard Requirements: <ul style="list-style-type: none"> · Delivery period 100 calendar days. · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier 			

· Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery

· The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market

Training Requirement

· The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility

· Certificate of training must be noted/signed by the head of Facility

After Sales Service

· Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model.

· List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available

Warranty Certificate

Warranty services for 1 year. Warranty includes labor and parts, free of charge. However, this excludes damages due to negligent acts of the user who shall be held liable for the cost of repair and spare parts. Only appointed service engineers shall be allowed to service the machine.

Special Provisions:

1. Non-removable DOH letters conspicuously displayed on the front

2. Property tagging (sticker) rewritable

*** **Nothing Follows** ***

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.

Technical Specifications

<p style="text-align: center;"> Republic of the Philippines Department of Health CENTER FOR HEALTH DEVELOPMENT CALABARZON </p>			
TECHNICAL SPECIFICATIONS			
Item No.:	10	Quantity	1
ABC :	P400,000.00	Country of Origin:	
Brand:		Model:	
PURCHASER'S SPECIFICATION		SUPPLIER'S SPECIFICATION	
<p style="text-align: center;">AUTOMATIC FILM PROCESSOR</p> <p> Power Supply: 220V Console Advance Imaging Plate ST-VI: 35 x 43 cm (14" x 17") ,x 35 x 35cm (14"x14"), 10"x12",8"x10" Min. 49 sec Up to 73 Ips/hr IP Cassete Type CC: 10 pixels / mm, 5 pixels /mm Min. 33 sec. Approx 165 sec. (Approx. 155 sec) * in case of 35 x 43cm (14" x 17") 1 10 Base T/100 Base TX 560 x 540 x 392 mm (22" x 21" x 15") 39 kg (86 lbs) Single phase 50-60 hz AC 120-240 V = 10% 1.9A (max) Operating Conditions: * Temperature: 15-30 C * Humidity : 15-80% RH (No dew condensation) * Atmosphere presure: 750-1060hPa Standard Requirements : · Delivery period 100 calendar days. · Certification to provide user's Manual in English Language a. Operations Manual 2 copies b. Service Manual 2 copies · Current and Valid Certificate of Manufacturer's compliance with ISO 13485 · Certificate of Commitment from the legal manufacturer of the unit that they will honor the warranty in case of change of distributor/supplier </p>			

<ul style="list-style-type: none"> · Transportation to site, delivery, installation and testing expenses on the site (health facility) for the account of the supplier and notification of the supplier to the recipient and respective regional offices prior to the delivery · The manufacturers certificate that the brand must be in the local market for at least five (5) years and 10 years in the international market <p>Training Requirement</p> <ul style="list-style-type: none"> · The supplier must provide training, orientation and product demonstration for users and maintenance personnel of the health facility · Certificate of training must be noted/signed by the head of Facility <p>After Sales Service</p> <ul style="list-style-type: none"> · Bidder's certificate that parts shall be available at the authorized Philippine service center/s for a period of five (5) years after the cessation of production of the same model. · List of service centers in major cities where equipment is installed and supplies / reagent outlet with contact landline numbers should be available <p>Warranty Certificate Minimum of three (3) years for both parts & service Warranty Certificate upon delivery, inspection and acceptance and after the provision of the required training of the end-user</p> <p>Special Provisions</p> <ol style="list-style-type: none"> 1. Non-removable DOH letters conspicuously displayed on the front 2. Property tagging (sticker) rewritable <p>*** Nothing Follows ***</p>	
--	--

NAME OF COMPANY

ADDRESS

SIGNATURE OVER PRINTED NAME
OF AUTHORIZED REPRESENTATIVE

TELEPHONE/ FAX NO.